


SprintIR™


High Speed Carbon Dioxide Sensor

SprintIR is a high speed (20 Hz) CO₂ sensor, ideally suited for applications which require capture of rapidly changing CO₂ concentrations including metabolic assessment and analytical instrumentation.

- High speed sensing (20Hz)
- Measurement ranges from 0 to 100%
- 3.3V supply
- Low power requirement 35mW
- Flow through adaptor (Optional)



SprintIR™ Sensor SprintIR™ Sensor
www.gassensing.co.uk
© GSS Ltd 2012




Specifications

CO ₂ Measurement	
Sensing Method	Non-dispersive infrared (NDIR) absorption Patented Gold-plated optics Patented Solid-state source and detector
Sample Method	Diffusion(Standard) / Flow through (with flow-through adaptor)
Measurement Range	0-5%, 0-20%, 0-60%, 0-100%
Accuracy	±70 ppm +/- 5% of reading ¹ (100% Range ±300 ppm +/-5% of reading ¹)
Measurement Noise	<10% of reading with no digital filtering
Non Linearity	< 1% of FS
Pressure Dependence	0.1% of reading per mbar in normal atmospheric conditions
Operating Pressure Range²	950 mbar to 10 bar ³

General Performance	
Warm-up Time	< 1 minute
Operating Conditions	0°C to 50°C (Standard) -25°C to 55°C (Extended range) 0 to 95% RH, non-condensing
Recommended Storage	-30°C to +70°C

Electrical/ Mechanical	
Power Input	<ul style="list-style-type: none"> • 3.2 to 5V. (3.3V recommended) • Peak current 100mA • Average Current <15mA
Power Consumption	35 mW
Output	UART only

Dimensions and Wiring Connections
2x5 0.1" header. Pin 1 is identified on the dimensional drawing.



Function	Pin #	Pin #	Function
0V	1	2	N/C
+3.3V	3	4	0V
Sensor Rx (in)	5	6	0V
Sensor Tx (out)	7	8	Zero N
N/C	9	10	Zero Air

Pin 2 should not be connected. Pins 4 and 6 do not require connection and are internally connected to GND.

The zeroing options are for hardware zeroing (both active low). These functions can also be implemented by sending a serial command (recommended).

Typical connections for digital interface are GND, 3.3V, Rx and Tx. Note that the V_h for the serial Tx line will be 3V regardless of the supply voltage.

Note 1: All measurements are at STP unless otherwise stated.

Note 2: Excludes Flow-through adapter. Contact GSS for more information

Note 3: External Pressure calibration required.

This documentation is provided on an as-is basis and no warranty as to its suitability or accuracy for any particular purpose is either made or implied. Gas Sensing Solutions Ltd will not accept any claim for damages howsoever arising as a result of use or failure of this information. Your statutory rights are not affected. This information is not intended for use in any medical appliance, device or system in which the failure of the product might reasonably be expected to result in personal injury. This document provides preliminary information that may be subject to change without notice.